

# BIOCHECK, INC.

323 Vintage Park Drive  
Foster City, CA 94404  
Tel: (650) 573-1968 Fax: (650) 573-1969

---

## PRODUCT SPECIFICATIONS

**Product Name:** T3 Standard/Sample Diluent

**Catalog Number:** MB-10102

**Lot Number:**

**Quantity:** \_\_\_\_\_ ml

**Product Form:** In human serum containing 1.0% (v/v) ProClin-300 as preservative.

**Storage:** At 2-8°C. **DO NOT FREEZE.**

**Suggested Use:** For use in the dilution of T3 Standards or patient samples in T3 EIA test.

**Viral Tests:** The source materials used for manufacture of this product tested negative for HBsAg, HIV 1&2 and HCV by FDA-approved methods. No methods can completely assure the absence of these agents and other infectious agents may be present.